

# CERTIFICATE OF ACCREDITATION

## MEDICINES CONTROL AUTHORITY OF ZIMBABWE

*Established by an Act of Parliament, the Medicines and Allied  
Substances Control amendment Act No 1 of 1996 Chapter 15:03*

**Facility Accreditation Number: TEST-11 0002**

is a SADCAS accredited Testing Laboratory  
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,  
Annexure "A", bearing the above accreditation number for

### PHYSICAL TESTING

*The facility is accredited in accordance with the recognized International Standard*

### ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation  
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and  
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States  
as a multi-economy accreditation body*

---

**Mrs Maureen P Mutasa**  
**SADCAS Chief Executive Officer**

**Date of Renewal of Accreditation: 12 February 2021**  
**Effective Date (Issue No: 1): 12 February 2021**  
**Certificate Expires: 11 February 2026**

**ANNEXURE A**  
**SCHEDULE OF ACCREDITATION**

**PHYSICAL TESTING**

**Laboratory Accreditation Number: TEST-11 0002 (ISO/IEC 17025:2017)**

<p><b><u>Permanent Address of Laboratory</u></b> Medicines Control Authority of Zimbabwe 106 Baines Avenue Harare Zimbabwe</p> <p><b><u>Postal Address</u></b> P O Box 10559 Harare Zimbabwe</p> <p><b><u>Tel</u></b> : +263 4 736 981/5 <b><u>Cell</u></b> : +263 77 214 5191 <b><u>Email</u></b> : <a href="mailto:mcaz@mcaz.co.zw">mcaz@mcaz.co.zw</a></p>		<p><b><u>Technical Signatories</u></b> : Mr T A Gonho (All Methods except Total Lubricants) Mr A Maqolo (All Methods) Ms L Nhukarume (All Methods) Ms Y R Zviuya (All Methods except Total Lubricants) Mr N O Shonhiwa (All Methods except Total Lubricants)</p> <p><b><u>Nominated Representative</u></b> : Ms A Chikowore</p> <p><b><u>Issue No</u></b> : 03 <b><u>Date of Issue</u></b> : 12 February 2021 <b><u>Expiry Date</u></b> : 11 February 2026</p>	
MATERIALS/PRODUCTS TESTED	TYPES OF TESTS/PROPERTIES MEASURED, RANGE OF MEASUREMENT	STANDARD SPECIFICATIONS, EQUIPMENT/TECHNIQUES USED	
<p><b><u>Male Rubber Latex Condoms</u></b></p>	<p>Dimension Testing</p> <ul style="list-style-type: none"> <li>• Determination of width</li> <li>• Determination of length</li> <li>• Determination of Thickness</li> </ul> <p>Freedom from holes</p> <p>Testing for package integrity</p> <p>Determination of Bursting Volume and Pressure</p>	<p>SOP MD-MT 028 / ISO 4074</p> <p>By Enersol automated width tester</p> <p>By Enersol automated length tester</p> <p>By Enersol automated thickness tester</p> <p>SOP MD-MT 030 / ISO 4074 by Enersol Ultra -fast electrical conductivity tester</p> <p>SOP MD-MT 031 / ISO 4074 by Enersol Package seal integrity tester</p> <p>SOP MD-MT 001; SOP MD-MT 003/ ISO 4074 BY Enersol air inflation testers</p>	

Original date of accreditation: 07 August 2015

Page 1 of 1

\_\_\_\_\_  
**Pinkie J Malebe**  
**SADCAS Technical Manager**