

CERTIFICATE OF ACCREDITATION

CALTES COMPANY LIMITED

Company Registration No: 141005006

Facility Accreditation Number: CAL-12 007

is a SADCAS accredited Calibration Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,
Annexure "A", bearing the above accreditation number for

PRESSURE METROLOGY

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States
as a multi-economy accreditation body*

Pinkie J Malebe
For SADCAS Chief Executive Officer

Effective Date (Issue No: 1): 14 June 2026
Certificate Expires: 30 August 2028

ANNEXURE A

SCHEDULE OF ACCREDITATION

PRESSURE METROLOGY

Laboratory Accreditation Number: CAL-12 007 (ISO/IEC 17025:2017)

<u>Permanent Address of Laboratory</u> Caltes Company Limited Mbezi Malamba Mawili – Kinyerezi Road Dar es Salaam Tanzania <u>Postal Address</u> P O Box 10051 Dar es Salaam Tanzania <u>Tel</u> : +255 75 440 8014 <u>Cell</u> : +255 71 999 2296 <u>Email</u> : b.shija@caltes.co.tz info@caltes.co.tz		<u>Technical Signatories</u> : Mr B M Shija (All items) <u>Nominated Representative</u> : Mr S H Bussara <u>Issue No</u> : 01 <u>Date of Issue</u> : 14 June 2026 <u>Expiry Date</u> : 30 August 2028		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ($\pm$)
1	Digital/Analogy Pressure Gauge (air medium)	CTC-PRG: 001	0 – 200 bars	At Caltes and On-site 0.04 %
2	Digital/Analogy Pressure Gauge (liquid medium)	CTC-PRG: 001	0 – 700 bars 800 – 1400 bars	0.03 % 0.03 %
3	Pressure Safety Valve	CTC-PSV: 001	0 – 600 bars	0.3 %
4	Pressure Transmitter Pressure Transducer through Electrical Simulation	CTC-CPT: 001	0 – 1000 bars (4 – 20) mA	0.01 % 0.003 mA

Original date of accreditation: 14 June 2026

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

Victor R Mundembe
For SADCAS Technical Manager