

CERTIFICATE OF ACCREDITATION

CALTES COMPANY LIMITED

Company Registration No: 141005006

Facility Accreditation Number: CAL-9 009

is a SADCAS accredited Calibration Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,
Annexure "A", bearing the above accreditation number for

VOLUME METROLOGY

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation
of a laboratory quality management system*

*SADCAS is a subsidiary organization of SADC. A memorandum of understanding between SADC and
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States
as a multi-economy accreditation body*

Pinkie J Malebe
For SADCAS Chief Executive Officer

Effective Date (Issue No: 1): 14 June 2026
Certificate Expires: 30 August 2028

ANNEXURE A

SCHEDULE OF ACCREDITATION

VOLUME METROLOGY

Laboratory Accreditation Number: **CAL-9 009 (ISO/IEC 17025:2017)**

Permanent Address of Laboratory Caltes Company Limited Mbezi Malamba Mawili – Kinyerezi Road Dar es Salaam Tanzania		Technical Signatories : Mr S H Bussara (All items)		
Postal Address P O Box 10051 Dar es Salaam Tanzania		Nominated Representative : Mr S H Bussara		
Tel : +255 75 440 8014 Cell : +255 78 332 6666 Email : s.bussara@caltes.co.tz ; info@caltes.co.tz ;		Issue No : 01 Date of Issue : 14 June 2026 Expiry Date : 30 August 2028		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)
2	VOLUME			At Caltes and On-site
2.2	Laboratory Glassware			
2.2.1	Flasks, Measuring Cylinders and Pycnometers	Internal: CTC – VOL – 002 & 003 Reference: SADCAS TR 19 R111-1, ISO 8106-2004 (E) EURAMET/cg – 18/v.04 ISO 8655 - 6	1 ml to 2000 ml	0,015% + 10 µl
2.2.2	Pipettes & Burettes		1 ml to 2000 ml	0,015% + 10 µl
2.2.3	Automated Pipettes		1 ml to 2000 ml	0,015% + 10 µl
2.2.4	Volume Dispensers		1 ml to 2000 ml	0,015% + 10 µl
2.3	Volume Dispensers			
2.3.1	Micropipettes Syringes	Internal: CTC – VOL - 001 Reference: SADCAS TR 19 R111-1, EURAMET/cg - 18/v.04 ISO 8655 - 6	10 µl to 200 µl 201 µl to 1000 µl 1001 µl to 10000 µl 10001 µl to 20000 µl	0,3 µl 2 µl 10 µl 30 µl

Original date of accreditation: 14 June 2026

Page 1 of 1

Victor R Mundembe
For SADCAS Technical Manager