

CERTIFICATE OF ACCREDITATION

MALAWI BUREAU OF STANDARDS (METROLOGY)

Established by Malawi Bureau of Standards Act (Cap 51:02)

Facility Accreditation Number: CAL-9 006

is a SADCAS accredited Calibration Laboratory
provided that all SADCAS conditions are complied with

This certificate is valid as per the scope stated in the accompanying schedule of accreditation,
Annexure "A", bearing the above accreditation number for

VOLUME METROLOGY

The facility is accredited in accordance with the recognized International Standard

ISO/IEC 17025:2017

*The accreditation demonstrates technical competency for a defined scope and the operation
of a laboratory quality management system*

*SADCAS is a subsidiarity organization of SADC. A memorandum of understanding between SADC and
SADCAS serves as the basis for the recognition of SADCAS by SADC Member States
as a multi-economy accreditation body*

Eve Christine Gadzikwa
SADCAS Chief Executive Officer

Date of Renewal of Accreditation 23 January 2024
Effective Date (Issue No: 1): 23 January 2024
Certificate Expires: 22 January 2029

ANNEXURE A

SCHEDULE OF ACCREDITATION

VOLUME METROLOGY

Laboratory Accreditation Number: **CAL-9 006 (ISO/IEC 17025:2017)**

<u>Permanent Address of Laboratory</u> Malawi Bureau of Standards Metrology Services Department Ali Hassan Mwinyi Road Blantyre Malawi		<u>Technical Signatories</u> : Mr R Makhole (All items) Mr M Banda (All items) Ms A P Khembo (All items) Mr I Lemison (All items)		
<u>Postal Address</u> P O Box 946 Blantyre Malawi		<u>Nominated Representative</u> : Dr S Chalimba		
<u>Tel</u> : +265 1870 488 <u>Cell</u> : +265 88 071 6329/99 278 5826 <u>Fax</u> : +265 1870 488 <u>Email</u> : stephencialimba@mbsmw.org agnesskhembo@mbsmw.org ;		<u>Deputy Nominated Representative</u> : Mr C Marorongwe		
		<u>Issue No</u> : 03 <u>Date of Issue</u> : 28 May 2026 <u>Expiry Date</u> : 22 January 2029		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	METHOD	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)
1	Micropipettes	Internal: MOP 005 Reference: SADCAS TR 19	50 µl - 100 µl 100 µl - 500 µl 500 µl - 1000 µl	0.2 µl 0.2 µl 0.2 µl
2	Glassware	Internal: MOP 005 Reference: SADCAS TR 19	1 ml-200ml 200 ml-1000 ml	0.5 ml 1.0 ml

Original date of accreditation: 23 January 2024

Page 1 of 1

Victor R Mundembe
 For SADCAS Technical Manager